

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/21/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.67: 0.079  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 1084 psi

BLANK TEST                    0.000   20:16  
INTERNAL STANDARD            ERROR   20:16

Internal Standard Error

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C            PASSED  
Breath Tube       = 45.9°C           PASSED

### PUMP INFO

Flow Rate = 4.814 L/M               PASSED

### DETECTOR INFO

PUMP ON                                PASSED  
PUMP OFF                                PASSED

### FILTER INFO

Filter 1                                 PASSED  
Filter 2                                 PASSED  
Filter 3                                 PASSED

### INTERNAL STANDARD

PASSED

File only. Instrument taken out of  
service on 02/21/2019 for  
multiple status messages.

JMV 03/18/19