

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100368

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/10/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.54: 0.079
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 1188 psi

BLANK TEST	0.000	21:31
INTERNAL STANDARD	VERIFIED	21:31
EXTERNAL STANDARD	0.080	21:31
BLANK TEST	0.000	21:32
EXTERNAL STANDARD	0.079	21:32
BLANK TEST	0.000	21:33
EXTERNAL STANDARD	0.079	21:34
BLANK TEST	0.000	21:34
EXTERNAL STANDARD	0.079	21:35
BLANK TEST	0.000	21:36
EXTERNAL STANDARD	0.079	21:36
BLANK TEST	0.000	21:37

Average = 0.0792
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 40.2°C PASSED

PUMP INFO

Flow Rate = 4.582 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

File Only: Instrument Removed from
service. BMB 11/11/19