

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form - DMCAF2011R1		
	Issuing Authority - Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100391

Date: 7/26/11

1. Wet Bath Simulator Solution

- a. Level 0.10
- b. Manufacturer Guth
- c. Lot 11050
- d. Expiration Date 4/6/13

2. Wet Bath Simulator

- a. Serial Number DR4177
- b. Date Calibrated 12/16/10

3. Attach Calibration Factors

Breath Alcohol Analyst   
 Charles Foster

1/2 CRF

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dnt: 100391  
-----

Date: 07/26/2011  
Time: 15:18:00

SUPERVISOR NAME:  
CHARLES FOSTER  
CERTIFICATE #: 7734

LOT #: 11050  
BOTTLE #: 481  
EXPIRATION: 04/06/2013

Ca	=	0.1000			
CAL	=	0.972426	0.800	<=	CAL < 1.200
b1	=	0.0001	0.0000	<=	b1 < 0.0040
b2	=	0.0047	0.0010	<=	b2 < 0.0100
b3	=	0.0000	0.0000	<=	b3 < 0.0040
Xq	=	0.0983	0.0500	<=	Xq < 0.2000
a21	=	1.231540	1.050	<=	a21 < 1.300
a31	=	0.487870	0.300	<=	a31 < 0.800

2/2 CFF

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100391\B-100391\_ECF-20140

2/19/2014