

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument  
State of Alaska*

Serial #: 100697

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/27/2022

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.76: 0.077  
LOT #: AG113402  
EXPIRATION: 05/14/2023  
TANK PRESSURE: 153 psi

BLANK TEST	0.000	04:31
INTERNAL STANDARD	VERIFIED	04:31
EXTERNAL STANDARD	INTERFERENCE	04:31

Blank Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 39.9°C	PASSED

### PUMP INFO

Flow Rate = 4.114 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED

File Only JMW 02/10/23