

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100374

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/12/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.17: 0.081
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 184 psi

BLANK TEST	0.000	21:04
INTERNAL STANDARD	VERIFIED	21:04
EXTERNAL STANDARD	0.077	21:04
BLANK TEST	0.000	21:05
EXTERNAL STANDARD	0.078	21:05
BLANK TEST	0.000	21:06
EXTERNAL STANDARD	0.078	21:06
BLANK TEST	0.000	21:07
EXTERNAL STANDARD	0.078	21:08
BLANK TEST	0.000	21:09
EXTERNAL STANDARD	INTERFERENCE	21:09

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 42.3°C	PASSED

PUMP INFO

Flow Rate = 4.336 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------