

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 7/8/16 Instrument Serial Number: 100682

Instrument Location: KAKE

Supervisor Name and Agency: TRP. UMBS, AST

Describe the Status Change or Issue with the Instrument:

RETURNED TO CRIME LAB. VOC NOT PERFORMED BECAUSE
DRUG GAS CALIBRATOR WAS EXPIRED.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 7/8/16 - 1000 HRS

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

O'Bryant, Colleen S (DPS)

Kake DMT o/s

From: O'Bryant, Colleen S (DPS)
Sent: Tuesday, April 26, 2016 12:40 PM
To: Edwards, John (DPS sponsored)
Cc: brandi.barnett@alaska.gov; Roth-Frey, Nicolette A (DPS); Foster, Charles R (DPS)
Subject: DMT

John,

We are going to mail you an empty case for your DMT, keyboard, power cord and breath hose. Could you please package up your DMT and mail it back to us.

Your agency is currently lacking a breath test supervisor, a dedicated phone line and the possibility of mold growing in the substation makes this environment unsuitable for the DMT to operate in. Thank you for all your help.

Colleen O'Bryant – Forensic Scientist III
Blood and Breath Alcohol
Alaska Scientific Crime Detection Laboratory
4805 Dr. Martin Luther King Jr. Ave
Anchorage, AK 99507

907-269-5783 (direct)
907-350-2111 (cell)