

In Field Instrument Review: 05/31/2018 – 07/17/2018

Prepared By Derek Walton

100355 (Seward PD): 1/19 RFI, 1/19 Interference Subject – No action taken

100360 (AST Soldotna): 4/25 RFI, 1/25 Internal standard error – No action taken

100363 (AST Anchor Point): 1/5 Pump error, 1/5 Standard out of range – Replacement tank sent on 8/14/18

100364 (APD): 1/85 RFI, 1/85 Pump error, 1/85 Undefined error – Returned to lab 7/12/18 due to fatal errors

100365 (North Pole PD): 2/14 Tank interference, 1/14 Internal standard error – Recommend monitoring

100366 (AST Nenana): 1/2 Standard out of range – Replacement tank sent on 8/14/18

100374 (APD): 1/15 Tank interference – No action taken

100378 (AST Fairbanks): 3/38 RFI – No action taken

100379 (Wasilla PD): 4/35 RFI, 1/35 Gas flow error – No action taken

100388 (AST Girdwood): 1/3 Pump error – No action taken

100389 (Cordova PD): 1/13 Blank error – No action taken

100395 (Hoonah PD): 1/8 Standard out of range – Blank error, Static analysis error, ambient fail on 7/18 –  
Instrument located near an external door that is regularly left open, door leads to vehicle parking area.  
Advised to close door during tests or relocate instrument.

100397 (North Slope Borough): 1/7 Blank error – No action taken

100402 (APD): 2/131 RFI, 2/131 Pump error, 1/131 Interference subject, 2/131 Blank error  
1/131 Ambient fail, 1/131 Undefined error – No action taken

100404 (St. Paul PD): 1/1 RFI – No action taken

100405 (Dillingham PD): 6/19 Tank interference – All on same day 6/13, training going on that day, successful VOC  
and breath tests afterwards

100413 (Homer PD): 2/19 RFI, 2/19 Interference subject – No action taken

100416 (Fairbanks PD): 1/53 RFI, 3/53 Pump error – No action taken

100422 (JBER): 1/7 RFI – No action taken

100423 (Ft Wainwright): 1/9 Blank error – No action taken

100666 (Bethel PD): 2/11 Internal standard error – Returned to lab on 7/27/18

100688 (Palmer PD): 4/46 RFI, 2/46 Interference subject – No action taken

100690 (Unalaska PD): 4/23 RFI – No action taken

100691 (Valdez PD): 1/11 RFI – No action taken

100698 (Nome PD): 1/12 Blank error – No action taken

100700 (Wrangell PD): 1/11 Pump error, 1/11 Blank error – No action taken

Does not include data from Eielson AFB, AST Aniak, or Galena PD



