

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 8 Dec 15 Instrument Serial Number: 100424

Instrument Location: 354 SFS 451 Divison st Eielson AFB AK

Supervisor Name and Agency: SSgt Jacobs, Joseph 354SFS/53T

Describe the Status Change or Issue with the Instrument:

RFI Detected when the room is clear of all Divices.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 8 Dec 15 / 12:28

For Use by SCDL

Additional Notes

Email completed form to [dipass@alaska.gov](mailto:dipass@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100424

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/08/2015

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.01  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C      PASSED  
Breath Tube      = 47.3°C      PASSED

RFI DETECTED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

\_\_\_\_\_  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this \_\_\_\_ day of \_\_\_\_\_, 20 \_\_\_\_