

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100355

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/13/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.04: 0.078
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 1153 psi

BLANK TEST	0.000	16:51
INTERNAL STANDARD	VERIFIED	16:52
EXTERNAL STANDARD	0.078	16:52
BLANK TEST	0.000	16:53
EXTERNAL STANDARD	0.077	16:53
BLANK TEST	0.000	16:54
EXTERNAL STANDARD	0.077	16:54
BLANK TEST	0.000	16:55
EXTERNAL STANDARD	0.076	16:55
BLANK TEST	0.000	16:56

RFI DETECTED

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 48.0°C	PASSED

PUMP INFO

Flow Rate = 3.776 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------