

# Alaska Scientific Crime Detection Laboratory

## DataMaster Calibration Form

Issued: 4/12/2018  
Effective: 4/13/2018

Version: DMCAF 2018 R0  
Status: Active

Instrument # 100688

Analyst Derck Walton

Date: 06/21/2019

### Reason for Calibration

Standard out of range, blank errors, internal standard error

#### 1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Auto
- c. Lot 18370
- d. Expiration Date 12/05/2020

#### 2. Wet Bath Simulator

- a. Serial Number SD1198
- b. Date Calibrated 03/06/2019

#### 3. Attach Calibration Factors

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100688  
-----

Date: 06/21/2019  
Time: 10:36:39

OPERATOR NAME:  
DEREK WALTON  
CERTIFICATE #: 9508

Ca = 0.1000  
CAL = 0.962316    0.800 <= CAL < 1.200  
b1 = 0.0004    0.0000 <= b1 < 0.0040  
b2 = 0.0041    0.0010 <= b2 < 0.0100  
b3 = 0.0009    0.0000 <= b3 < 0.0040  
Xq = 0.0995    0.0500 <= Xq < 0.2000  
a21 = 1.138785    1.050 <= a21 < 1.300  
a31 = 0.501704    0.300 <= a31 < 0.800