

	State of Alaska Scientific Crime Detection Laboratory	
	DataMaster Calibration Form - DMCAF2011R1	
	Issuing Authority - Nita Bolz, Scientific Director	
Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100381

Date: 4/26/11

1. Wet Bath Simulator Solution

- a. Level .100
- b. Manufacturer Guth
- c. Lot 10260
- d. Expiration Date 10/25/12

2. Wet Bath Simulator

- a. Serial Number DR-4161
- b. Date Calibrated 8/22/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. O'Bryant

COS/p

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dnt: 100381

Date: 04/26/2011
Time: 08:49:42

SUPERVISOR NAME:
COLLEEN S O'BRYANT
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 10260
EXPIRATION: 10/25/2012

Ca	=	0.1000		
CAL	=	0.974059	0.800	<= CAL < 1.200
b1	=	0.0000	0.0000	<= b1 < 0.0040
b2	=	0.0029	0.0010	<= b2 < 0.0100
b3	=	0.0000	0.0000	<= b3 < 0.0040
Xq	=	0.0904	0.0500	<= Xq < 0.2000
a21	=	1.090518	1.050	<= a21 < 1.300
a31	=	0.416583	0.300	<= a31 < 0.600

COB/6

Final page for report

\

VANCCLSAN01\JTRAX\CASEDOCS\B-100381\B-100381_ECF-20140

7/3/2014