

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100375

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 11/17/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 27.19: 0.073  
LOT #: AG826101  
EXPIRATION: 09/18/2020  
TANK PRESSURE: 1177 psi

BLANK TEST	0.000	14:00
INTERNAL STANDARD	VERIFIED	14:00
EXTERNAL STANDARD	0.073	14:00
BLANK TEST	0.000	14:01
EXTERNAL STANDARD	INTERFERENCE	14:01

Blank Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 46.0°C	PASSED

### PUMP INFO

Flow Rate = 4.308 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------