

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100365

Date: 9/22/15

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Software manage pop-ups - unable to run tests

2. Visual Inspection

3. Troubleshooting/Repair

unable to perform any troubleshooting due to
pop-up boxes (see attached)

Attempted to upload - unsuccessful

Attempted to connect in interactive mode -
unsuccessful

Returning to NPAS

Last upload prior to instrument being taken out
of service was 9/4/15

Out of service date was 9/11/15

4. VOC (attached) N/A

5. Internal Standard value $\leq 3.00\%$ Yes No N/A

6. Verify software version N/A

Return to Factory for Repair

Ready For Calibration

Ready For Certification

```
System.NullReferenceException:  
NullReferenceException  
at  
NPAS.TicketDataBase.GetDownloadTickets()  
at ModemClass.Modem.ModemLoop()
```

Lot Number

Expiration Date

T
Ch

Cc

