

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100357

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/05/2020

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.89: 0.080  
LOT #: AG923401  
EXPIRATION: 08/22/2021  
TANK PRESSURE: 1100 psi

BLANK TEST                      0.000    11:39  
INTERNAL STANDARD            ERROR    11:39

Internal Standard Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C            PASSED  
Breath Tube     = 44.1°C            PASSED

### PUMP INFO

Flow Rate = 4.549 L/M              PASSED

### DETECTOR INFO

PUMP ON                              PASSED  
PUMP OFF                              PASSED

### FILTER INFO

Filter 1                              PASSED  
Filter 2                              PASSED  
Filter 3                              PASSED

### INTERNAL STANDARD

PASSED

File only DMW 06/09/20