

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 09-21-16 Instrument Serial Number: 100689

Instrument Location: Kotzebue Police Department

Supervisor Name and Agency: Ofc. James Rea, Kotzebue Police Dept.

Describe the Status Change or Issue with the Instrument:

Screen froze, attempted to restart instrument and it would not
come back on. Checked all power connection and checked plug in
other surge protectors and still no power.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time:

Out of Service Date/Time: 09-21-16 @ 0430

| For Use by SCDL | Additional Notes |
|-----------------|------------------|
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740