

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 4/12/2018 Instrument Serial Number: 100376

Instrument Location: UNALASKA DEPT. OF Public Safety.

Supervisor Name and Agency: KEVIN WOOD UNALASKA

Describe the Status Change or Issue with the Instrument:

DATAMASTER SN 100376 had a Hard Disk error  
which was unable to be cleared. The ALASKA SCIENTIFIC  
CRIME DETECTION LABORATORY PERSONNEL AS STATED they  
were having a difficult time communicating with the  
DATAMASTER, and was unable to retrieve necessary  
OPERATION Files VOCs.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 4/12/2018 <sup>1200HR.</sup> SERIAL NUMBER 100376

For Use by SCDL

Additional Notes

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740