

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 12/27/2023

Instrument #: **100358**

Instrument Location: AST/Mat-Su West

Supervisor Name and Agency: Colleen O'Bryant, SCDL

Describe the Status Change or Issue with the Instrument:

I reached out to AST Breath Test Supervisors in the valley to have them remind local operators about RFI in the vicinity of the instrument. See attached email for more information. COB 2023.12.27

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: N/A

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

O'Bryant, Colleen S (DPS)

From: O'Bryant, Colleen S (DPS)
Sent: Wednesday, December 27, 2023 11:29 AM
To: Barela, Mariano D (DPS); Phillips, Lee E (DPS); Crockett, Jason A (DPS)
Subject: RFI Status Messages on MSW DataMaster DMT

AST/Valley Breath Test Supervisors,

I have reviewed the tests from 100358 from the last two verification of calibration cycles. During September 6th – October 24rd – there were 7 RFI status messages out of 20 tests total. During October 25th – December 12th – there were 5 RFI status messages out of 19 tests total. This is a higher than normal amount of status messages. Please speak with your operators and remind them to use precautions and awareness when using radios in the vicinity of the DataMaster. I know the placement of the MSW DMT in the squad room can be difficult to isolate RFI.

Let me know if you have any questions.

Colleen O'Bryant

**Forensic Scientist III - Chemistry
Alaska Scientific Crime Detection Laboratory
4805 Dr. Martin Luther King Jr. Ave
Anchorage, Alaska 99507**

**907-269-5783 (direct)
907-350-2111 (cell)**