



State of Alaska Scientific Crime Detection Laboratory

DataMaster Acceptance Form – DMAF2011R0

Issuing Authority – Nita Bolz, Scientific Director

Approval date: 03/01/11

Effective date: 03/01/11

Archive date: Active

DataMaster DMT Acceptance Form

Instrument # 100357

Date: 3/1/11

Analyst: Colleen O'Bryant

- 1. Check in instrument in to JusticeTrax Breath Alcohol Database
- 2. Visual Inspection
- 3. Update software
- 4. Print calibration factors (attached)
- 5. NPAS calibration paperwork (attached)
- 6. Barcode Instrument
- 7. State of Alaska Property ID# 10140666

**B-100357** 100357



QSWNV35T0ITEWL

100357


NPAS

CALIBRATION FACTORS

Alaska Department of Public Safety  
DALAMASTER dnt: 103357

Date: 01/12/2011  
Time: 07:00:53

Ca = 0.1000  
CAL = 0.908346 0.800 <= CAT < 1.200  
b1 = 0.0000 0.0000 <= b1 < 0.0040  
b2 = 0.0046 0.0015 <= b2 < 0.0100  
b3 = 0.0001 0.0000 <= b3 < 0.0040  
Xq = 0.0952 0.0500 <= Xq < 0.2000  
a21 = 1.185905 1.000 <= a21 < 1.300  
a31 = 0.533813 0.300 <= a31 < 0.800



**National Patent  
Analytical Systems, Inc.**

*Certificate of Calibration*

100357

Instrument Serial Number

10180

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

1-12-11

Date

Charles E. Saunders

Technician

Final page for report

\

VANCCLSAN01\JTRAX\CASEDOCS\B-100357\B-100357\_ECF-20140

6/27/2014