

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100684

Received From: DPS Academy

Date: 12/14/22

Analyst: C. O Bryant

Reason for instrument return:

undefined error
unable to load ticket db

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload unable to upload due to ticket db pop up*

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)

* Successfully uploaded during repair + checked for VOCs.
CAS 12/14/22