

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100684

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/27/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.20: 0.075
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 369 psi

BLANK TEST	0.000	04:43
INTERNAL STANDARD	VERIFIED	04:43
EXTERNAL STANDARD	0.071	04:44
BLANK TEST	0.000	04:45
EXTERNAL STANDARD	0.070	04:45
BLANK TEST	0.000	04:46
EXTERNAL STANDARD	0.071	04:46
BLANK TEST	0.000	04:47
EXTERNAL STANDARD	0.072	04:47
BLANK TEST	0.000	04:48
EXTERNAL STANDARD	0.072	04:48
BLANK TEST	0.000	04:49

Average = 0.0712
Std Dev = 0.0008

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO

Flow Rate = 3.842 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only : Instrument being removed from
service. BMB 3/8/18