	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Acceptance Form - DMAF2011R0		
	Issuing Authority - Nita Bolz, Scientific Director		
Approval date: 03/01/11	Effective date: 03/01/11	Archive date: Active	

DataMaster DMT Acceptance Form

Instrument # 100668

Date: 10/12/11

Analyst: CTCF

1. Check in instrument in to JusticeTrax Breath Alcohol Database
2. Visual Inspection
3. Update software
4. Print calibration factors (attached)
5. NPAS calibration paperwork (attached)
6. Barcode Instrument
7. State of Alaska Property ID# 12-F0356

**B-100668** 100668




QSHS03C20UZTEN B-100668  
Breath Program

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100668  
-----

Date: 09/21/2011  
Time: 14:00:28

Ca = 0.1500  
CAL = 0.988351 0.800 <= CAL < 1.200  
b1 = 0.0000 0.0000 <= b1 < 0.0040  
b2 = 0.0016 0.0010 <= b2 < 0.0100  
b3 = 0.0000 0.0000 <= b3 < 0.0040  
Xq = 0.0947 0.0500 <= Xq < 0.2000  
a21 = 1.169724 1.050 <= a21 < 1.300  
a31 = 0.527723 0.300 <= a31 < 0.800



**National Patent  
Analytical Systems, Inc.**

*Certificate of Calibration*

100668

Instrument Serial Number

11030

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

9/22/11

Date

Charles E. Lawrence

Technician

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100668\B-100668 ECF-20140

7/10/2014