

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 5/29/20 Instrument Serial Number: 100345

Instrument Location: VALDEZ POLICE DEPARTMENT

Supervisor Name and Agency: SGT. CHAO CLEMENTS VPD

Describe the Status Change or Issue with the Instrument:

DISCOVERED INSTRUMENT ONLY RAN ONE EXTERNAL STANDARD ON
SUBJECT TEST. I CONTACTED THE CRIME LAB AND WAS ABLE TO
GET THE ISSUE FIXED REMOTELY. I THEN VERIFIED THE TEST
SEQUENCE WAS CORRECTED.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740