

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 1/8/2025

Instrument #: 100414

Instrument Location: Glacier Bay National Park

Supervisor Name and Agency: Jarred Mitrea – National Park Service

Describe the Status Change or Issue with the Instrument:

Instrument was showing errors with it diagnostic check and was removed from service to be sent to the State Crime lab.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: January 1, 2025 @ 1351 hours

For Use by SCDL

Additional Notes

Internal standard errors DJW 01/08/2025

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor