

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 04/22/2019 Instrument Serial Number: 100415

Instrument Location: 304 Lake Street, Sitka, AK 99835

Supervisor Name and Agency: Sgt. Isaiah Clark, Sitka Police Department

Describe the Status Change or Issue with the Instrument:

Forensic Scientist Derek Walton contacted me, in regards to our current Datamaster experiance internal standard and blank errors.
I received a new Datamaster from the SCDL.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 04/22/2019/2213

For Use by SCDL	Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740