

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 6/10/16 Instrument Serial Number: 1 00366  
Instrument Location: Wildwood Pretrial Facility - Kenai  
Supervisor Name and Agency: Ryan Grimm - Kenai P.D.

Describe the Status Change or Issue with the Instrument:

was not performing VOC's on own w/out  
tank change.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 6/10/16 2215

For Use by SCDL

Additional Notes

VOCs were failing due to chopper frequency on XML A1c  
of VOC. Was getting diagnostic fail and every thing read  
passed.

Email completed form to \_\_\_\_\_

For questions contact the Breath Alcohol Program at 907-269-5740