

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMCIF 2015 R1
Status: Active

Instrument S/N 100696

Received From Hooper Bay RD

Date 10/15/15

Analyst Colleen O'Bryant

1. Reason for instrument return:

Interference detected on external standard and internal standard error on 9/5/15 VOC.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From NPAS

7. NPAS Repair Paperwork (attached)