

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100375

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 11/07/2018

## External Standard Test Values

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C PASSED  
Breath Tube = 44.6°C PASSED

### PUMP INFO

Flow Rate = 4.459 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF FAILED

### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD

INTERNAL STANDARD FAILED

Diagnostic Check Failed