

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 09-14-2021 Instrument Serial Number: 100369

Instrument Location: AST Bethel Post

Supervisor Name and Agency: Tatiana Lee - ASCDL

Describe the Status Change or Issue with the Instrument:

AST Bethel Post's DMT did not populate a VOC on 09-05-2021. I called the post and spoke with Trooper Scott Corino who confirmed that the expired tank is still installed in the instrument. Trooper Corino's certification expired on 12-31-2020 and the only officer with a current certification is out on leave. I walked Trooper Corino through changing a tank. A VOC was initiated so I advised him to send me a copy of the VOC via email and to restart the DMT after the VOC has been completed.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740