


| | | | |
|---|---|----------------------|--|
|  | State of Alaska Scientific Crime Detection Laboratory | | |
| | DataMaster Calibration Form -- DMCAF2011R1 | | |
| | Issuing Authority -- Nita Bolz, Scientific Director | | |
| Approval date: 03/10/11 | Effective date: 03/10/11 | Archive date: Active | |

DataMaster DMT Calibration Form

Instrument # 100379Date: 3/22/11

1. Wet Bath Simulator Solution

- a. Level 100
 b. Manufacturer Bott
 c. Lot ~~9340~~ 9380
 d. Expiration Date 12/17/11

2. Wet Bath Simulator

- a. Serial Number DR4161
 b. Date Calibrated 9/22/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. O'Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dnt: 100379

Date: 03/22/2011
Time: 12:35:43

SUPERVISOR NAME:
COLLEEN S O'BRYANT
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 9380
EXPIRATION: 12/17/2011

| | | | | | |
|-----|---|----------|--------|----|-------------|
| Ca | = | 0.1000 | | | |
| CAL | = | 0.994139 | 0.800 | <= | CAL < 1.200 |
| b1 | = | 0.0005 | 0.0000 | <= | b1 < 0.0040 |
| b2 | = | 0.0044 | 0.0010 | <= | b2 < 0.0100 |
| b3 | = | 0.0002 | 0.0000 | <= | b3 < 0.0040 |
| Xq | = | 0.0812 | 0.0500 | <= | Xq < 0.2000 |
| a21 | = | 1.116438 | 1.050 | <= | a21 < 1.300 |
| a31 | = | 0.451907 | 0.300 | <= | a31 < 0.800 |

COB

Final page for report

\

VANCSCLSAN01\JTRAX\CASEDOCS\B-100379\B-100379_ECF-20140

7/3/2014