

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100382

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/06/2020

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.55: 0.079  
LOT #: AG826101  
EXPIRATION: 09/18/2020  
TANK PRESSURE: 596 psi

BLANK TEST	0.000	19:44
INTERNAL STANDARD	VERIFIED	19:44
EXTERNAL STANDARD	INTERFERENCE	19:44

Interference Detected

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 48.0°C	PASSED

### PUMP INFO

Flow Rate = 5.017 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only JMW 06/04/20