

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 2/9/2023

Instrument #: 100667

Instrument Location: Anchorage Jail – South processing room

Supervisor Name and Agency: Jason Cusack, Anchorage Police

Describe the Status Change or Issue with the Instrument:

I replaced instrument 100668 with this instrument in the South OUI processing room at the Anchorage Jail.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 2/9/2023 2355 hours

Out of Service Date/Time: [Click or tap here to enter text.](#)

For Use by SCDL

Additional Notes

Sent as a replacement DMT. Old DMT (100668) was pulled for filter wheel errors. TLM 02/10/2023

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor