

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100683

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 06/26/2020

## External Standard Test Values

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 41.1°C PASSED

### PUMP INFO

Flow Rate = 3.986 L/M PASSED

### DETECTOR INFO

PUMP ON FAILED  
PUMP OFF PASSED

### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD

PASSED

Diagnostic Check Failed