

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100675

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/26/2025

External Standard Test Values

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 46.2°C	PASSED

PUMP INFO	
Flow Rate = 4.920 L/M	PASSED

DETECTOR INFO	
PUMP ON	FAILED
PUMP OFF	FAILED

FILTER INFO	
Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

Diagnostic Check Failed

File only
CWS-3/6/25