



State of Alaska Scientific Crime Detection Laboratory

DataMaster Acceptance Form - DMAF2011R0

Issuing Authority - Nita Bolz, Scientific Director

Approval date: 03/01/11 Effective date: 03/01/11 Archive date: Active

DataMaster DMT Acceptance Form

Instrument # 100359

Date: 3/2/11

Analyst: Colleen O'Bryen


1. Check in instrument in to JusticeTrax Breath Alcohol Database
2. Visual Inspection
3. Update software
4. Print calibration factors (attached)
5. NPAS calibration paperwork (attached)
6. Barcode Instrument
7. State of Alaska Property ID# 10140669

B-100359 100359



OSMNV35U0K16TX
NPAS

100359



National Patent

Analytical Systems, Inc.

Certificate of Calibration

100359

Instrument Serial Number

10180

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

1-12-11

Date

Charles E. Townsend

Technician

CALIBRATION FACTORS

Alaska Department of Public Safety
LALAMASTER dat: 100359

Date: 01/12/2011
Time: 12:10:27

Ca = 0.1000
CAL = 0.900148 0.800 <= CAL < 1.200
b1 = 0.0001 0.0000 <= b1 < 0.0040
b2 = 0.0034 0.0010 <= b2 < 0.0100
b3 = 0.0000 0.0000 <= b3 < 0.0040
Xq = 0.0901 0.0500 <= Xq < 0.2500
a21 = 1.152894 1.050 <= a21 < 1.300
a31 = 0.403409 0.300 <= a31 < 0.500

Final page for report

\

VANCCLSAN01\JTRAX\CASEDOCS\B-100359\B-100359_ECF-20140

6/27/2014