

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100412


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/24/2023

External Standard Test Values	Diagnostic Check																																																																								
<p>EXTERNAL STANDARD INFORMATION</p> <p>NOMINAL: 0.080</p> <p>TARGET AT 30.48: 0.082</p> <p>LOT #: AG310901</p> <p>EXPIRATION: 04/19/2025</p> <p>TANK PRESSURE: 1077 psi</p>	<p>VERSIONS</p> <p>DMT: 3.02</p> <p>PIC: 3.02</p> <p>Modem: 2.6</p> <p>Questions: 2.2</p>																																																																								
<table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 20%;">BLANK TEST</td> <td style="width: 10%;">0.000</td> <td style="width: 10%;">12:02</td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> </tr> <tr> <td>INTERNAL STANDARD</td> <td>VERIFIED</td> <td>12:02</td> <td></td> <td></td> <td></td> </tr> <tr> <td>EXTERNAL STANDARD</td> <td>0.081</td> <td>12:02</td> <td></td> <td></td> <td></td> </tr> <tr> <td>BLANK TEST</td> <td>0.000</td> <td>12:03</td> <td></td> <td></td> <td></td> </tr> <tr> <td>EXTERNAL STANDARD</td> <td>0.081</td> <td>12:03</td> <td></td> <td></td> <td></td> </tr> <tr> <td>BLANK TEST</td> <td>0.000</td> <td>12:04</td> <td></td> <td></td> <td></td> </tr> <tr> <td>EXTERNAL STANDARD</td> <td>0.081</td> <td>12:04</td> <td></td> <td></td> <td></td> </tr> <tr> <td>BLANK TEST</td> <td>0.000</td> <td>12:05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>EXTERNAL STANDARD</td> <td>0.081</td> <td>12:05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>BLANK TEST</td> <td>0.000</td> <td>12:06</td> <td></td> <td></td> <td></td> </tr> <tr> <td>EXTERNAL STANDARD</td> <td>0.081</td> <td>12:06</td> <td></td> <td></td> <td></td> </tr> <tr> <td>BLANK TEST</td> <td>0.000</td> <td>12:07</td> <td></td> <td></td> <td></td> </tr> </table>	BLANK TEST	0.000	12:02				INTERNAL STANDARD	VERIFIED	12:02				EXTERNAL STANDARD	0.081	12:02				BLANK TEST	0.000	12:03				EXTERNAL STANDARD	0.081	12:03				BLANK TEST	0.000	12:04				EXTERNAL STANDARD	0.081	12:04				BLANK TEST	0.000	12:05				EXTERNAL STANDARD	0.081	12:05				BLANK TEST	0.000	12:06				EXTERNAL STANDARD	0.081	12:06				BLANK TEST	0.000	12:07				<p>TEMPERATURES</p> <p>Sample Chamber = 48.9°C PASSED</p> <p>Breath Tube = 48.1°C PASSED</p> <p>PUMP INFO</p> <p>Flow Rate = 4.897 L/M PASSED</p> <p>DETECTOR INFO</p> <p>PUMP ON PASSED</p> <p>PUMP OFF PASSED</p> <p>FILTER INFO</p> <p>Filter 1 PASSED</p> <p>Filter 2 PASSED</p> <p>Filter 3 PASSED</p> <p>INTERNAL STANDARD PASSED</p>
BLANK TEST	0.000	12:02																																																																							
INTERNAL STANDARD	VERIFIED	12:02																																																																							
EXTERNAL STANDARD	0.081	12:02																																																																							
BLANK TEST	0.000	12:03																																																																							
EXTERNAL STANDARD	0.081	12:03																																																																							
BLANK TEST	0.000	12:04																																																																							
EXTERNAL STANDARD	0.081	12:04																																																																							
BLANK TEST	0.000	12:05																																																																							
EXTERNAL STANDARD	0.081	12:05																																																																							
BLANK TEST	0.000	12:06																																																																							
EXTERNAL STANDARD	0.081	12:06																																																																							
BLANK TEST	0.000	12:07																																																																							
<p>Average = 0.0810</p> <p>Std Dev = 0.0000</p>																																																																									

I, Derek J. Walton, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) As of the date signed below, I certify that the calibration of the referenced instrument was accurate on the date in which the above tests were performed and therefore certified for evidentiary use in the State of Alaska.



 Derek J. Walton
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 21 day of 11, 20 23



 Notary Public
 My Commission Expires With Office



OK
 11/5/23