

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 05/15/2020
Effective: 05/15/2020

Version: DMEF 2020 R0
Status: Active

Instrument # 100365

Start Date: 10/28/2021

Analysts: Carlic Plaister

Reason for instrument evaluation

Undefined error followed by unsaved breath tests

Visual Inspection

- | | |
|---|---|
| <input checked="" type="checkbox"/> Clean | <input checked="" type="checkbox"/> Replace missing screws |
| <input checked="" type="checkbox"/> Cermetex Box is secure | <input checked="" type="checkbox"/> Ensure modem cable present |
| <input checked="" type="checkbox"/> Gas Valve upright or secured to wall | <input checked="" type="checkbox"/> All tubing and connections are secure |
| <input checked="" type="checkbox"/> O-ring inspection and ensure 2 O-rings in replacement bag | |

Troubleshooting/Repair

Replaced o-ring. Noted some external damage on the top left of the outer casing (small gouge in the metal).
VOC passed. VOC & Breath test appeared normal & saved appropriately.
Reuploaded 3.02 DM7 software.
On 2nd VOC internal standard during diagnostic above 3%.
Repeated breath test. VOC & breath test saved appropriately.
Recommend Calibration.

- | | |
|---|---|
| <input checked="" type="checkbox"/> VOC (attached) | |
| <input checked="" type="checkbox"/> Internal Standard value $\leq 3.00\%$ | * passed on first VOC, failed on second VOC ckg |
| <input checked="" type="checkbox"/> Verify software version | 10/29/2021 |

- | | |
|---|---|
| <input type="checkbox"/> Return to Factory for Repair | <input checked="" type="checkbox"/> Ready for Calibration |
| <input type="checkbox"/> Ready for Certification | <input type="checkbox"/> Proceed to Close Out |

End Date: 10/29/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/28/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.32: 0.078
LOT #: AG018102
EXPIRATION: 06/29/2022
TANK PRESSURE: 385 psi

BLANK TEST	0.000	14:25
INTERNAL STANDARD	✓ VERIFIED	14:26
Xq = 0.103 (1.77%)		
EXTERNAL STANDARD	0.077	14:26
X[1] = 0.0773 (-0.0015) (-0.0017)		
BLANK TEST	0.000	14:27
EXTERNAL STANDARD	0.077	14:27
X[1] = 0.0774 (-0.0017) (-0.0019)		
BLANK TEST	0.000	14:28
EXTERNAL STANDARD	0.078	14:28
X[1] = 0.0776 (-0.0017) (-0.0017)		
BLANK TEST	0.000	14:29
EXTERNAL STANDARD	0.077	14:29
X[1] = 0.0773 (-0.0019) (-0.0018)		
BLANK TEST	0.000	14:30
EXTERNAL STANDARD	0.077	14:30
X[1] = 0.0769 (-0.0023) (-0.0017)		
BLANK TEST	0.000	14:31

Average = 0.0772
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 43.5°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.83 V	PASSED
(1.10 V - 2.56 V)	
Cooler Voltage = 1.84 V	PASSED
(1.10 V - 2.58 V)	
Bias Voltage = 80 V	PASSED
(48 V - 112 V)	
Chopper Freq = 514 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.3 in	

PUMP INFO

Flow Rate = 4.605 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.005 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	✓	PASSED
Xq = 0.103	1.60%	(0.00% - 4.00%)

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100365

Date: 10/28/2021
Time: 14:43:17

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX


OPERATOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436

DEPT/AGENCY:
TEST
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.33: 0.078
LOT #: AG018102
EXPIRATION: 06/29/2022

DIAGNOSTIC CHECK	PASSED	14:44
BLANK TEST	0.000	14:44
INTERNAL STANDARD	VERIFIED	14:44
Xg = 0.103 (1.60%)		
EXTERNAL STANDARD	0.078	14:44
X[1] = 0.0777 (-0.0014)	(0.0007)	
BLANK TEST	0.000	14:45
SUBJECT SAMPLE (Vol=2.71L)	0.000	14:46
X[1] = 0.0000 (-0.0023)	(-0.0012)	
BLANK TEST	0.000	14:47
EXTERNAL STANDARD	0.077	14:47
X[1] = 0.0774 (-0.0017)	(-0.0008)	
BLANK TEST	0.000	14:48

OKG



VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/29/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.54: 0.079
LOT #: AG018102
EXPIRATION: 06/29/2022
TANK PRESSURE: 367 psi

BLANK TEST	0.000	09:04
INTERNAL STANDARD	✓ VERIFIED	09:04
Xq = 0.102 (2.27%)		
EXTERNAL STANDARD	0.077	09:04
X[1] = 0.0768 (-0.0010) (0.0017)		
BLANK TEST	0.000	09:05
EXTERNAL STANDARD	0.077	09:05
X[1] = 0.0767 (-0.0022) (0.0011)		
BLANK TEST	0.000	09:06
EXTERNAL STANDARD	0.077	09:07
X[1] = 0.0771 (-0.0015) (0.0017)		
BLANK TEST	0.000	09:07
EXTERNAL STANDARD	0.078	09:08
X[1] = 0.0775 (-0.0011) (0.0014)		
BLANK TEST	0.000	09:09
EXTERNAL STANDARD	0.077	09:09
X[1] = 0.0772 (-0.0018) (0.0004)		
BLANK TEST	0.000	09:10

Average = 0.0772
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 44.1°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.83 V	PASSED
(1.10 V - 2.56 V)	
Cooler Voltage = 1.84 V	PASSED
(1.10 V - 2.58 V)	
Bias Voltage = 80 V	PASSED
(48 V - 112 V)	
Chopper Freq = 523 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.5 in	

PUMP INFO

Flow Rate = 4.583 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.003 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD PASSED

Xq = 0.101 3.11% (0.00% - 4.00%)

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100365

Date: 10/29/2021
Time: 09:12:59

SUBJECT NAME:

TEST TEST
LICENSE #: XXXX

OPERATOR NAME:

CARLIE GLAISTER
OPERATOR'S #: 10436

DEPT/AGENCY:

TEST
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.54: 0.079
LOT #: AG018102
EXPIRATION: 06/29/2022

DIAGNOSTIC CHECK	PASSED	09:14
BLANK TEST	0.000	09:14
INTERNAL STANDARD	VERIFIED	09:14
$X_q = 0.102$ (2.44%)		
EXTERNAL STANDARD	0.077	09:15
$X[1] = 0.0772$ (-0.0016)	(0.0025)	
BLANK TEST	0.000	09:15
SUBJECT SAMPLE (Vol=2.30L)	0.000	09:16
$X[1] = 0.0000$ (-0.0010)	(0.0000)	
BLANK TEST	0.000	09:17
EXTERNAL STANDARD	0.078	09:17
$X[1] = 0.0779$ (-0.0006)	(0.0018)	
BLANK TEST	0.000	09:18

CXG

Alcohol (g/210L)
Flow Rate (L/M)
Min. Flow Rate (L/M)

