

# Alaska Scientific Crime Detection Laboratory

## DataMaster Calibration Form

Issued: 4/12/2018  
Effective: 4/13/2018

Version: DMCAF 2018 R0  
Status: Active

Instrument # 100402

Analyst Brandi Barnett  
& Carlie Glaister

Date: 10/12/21

### Reason for Calibration

Recalibrate due to refurbishment at factory.

#### 1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Lifecloc
- c. Lot 21DP41158
- d. Expiration Date 4/29/2023

#### 2. Wet Bath Simulator

- a. Serial Number D125298
- b. Date Calibrated 6/16/2021

#### 3. Attach Calibration Factors

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100402  
-----

Date: 10/12/2021  
Time: 14:30:02

OPERATOR NAME:  
CARLIE GLAISTER  
CERTIFICATE #: 10436

*BMB*  
*10/12/21*

Ca = 0.1000  
CAL = 0.953874    0.800 <= CAL < 1.200  
b1 = 0.0010    0.0000 <= b1 < 0.0040  
b2 = 0.0066    0.0010 <= b2 < 0.0100  
b3 = 0.0008    0.0000 <= b3 < 0.0040  
Xq = 0.0759    0.0500 <= Xq < 0.2000  
a21 = 1.176229    1.050 <= a21 < 1.300  
a31 = 0.376163    0.300 <= a31 < 0.800