

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 16 Aug 19 Instrument Serial Number: 100415

Instrument Location: Ft Wainwright

Supervisor Name and Agency: Rossignol, Corey FWLWD

Describe the Status Change or Issue with the Instrument:

New replacement instrument

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 16 Aug 19 0905

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740