

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 08/08/2017
Effective: 08/10/2017

Version: DMEF2017 R1
Status: Active

Instrument # 100387

Date: 1/19/18

Analyst: Brandi Barnett

1. Reason for instrument evaluation

Preventative Maintenance

2. Visual Inspection (internal and external)

3. O-ring inspection

4. Troubleshooting/Repair

none

5. VOC (attached)

6. Internal Standard value $\leq 3.00\%$ Yes No

7. Verify software version

Return to Factory for Repair

Ready for Calibration

Ready for Certification

Proceed to Close Out

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100387

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/19/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.90: 0.077
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 179 psi

BLANK TEST	0.000	14:50
INTERNAL STANDARD	VERIFIED	14:50
Xq = 0.098 (0.18%)		
EXTERNAL STANDARD	0.079	14:51
X[1] = 0.0790 (-0.0017) (-0.0019)		
BLANK TEST	0.000	14:51
EXTERNAL STANDARD	0.079	14:52
X[1] = 0.0786 (-0.0014) (-0.0018)		
BLANK TEST	0.000	14:53
EXTERNAL STANDARD	0.079	14:53
X[1] = 0.0790 (-0.0017) (-0.0014)		
BLANK TEST	0.000	14:54
EXTERNAL STANDARD	0.079	14:54
X[1] = 0.0788 (-0.0013) (-0.0010)		
BLANK TEST	0.000	14:55
EXTERNAL STANDARD	0.078	14:55
X[1] = 0.0778 (-0.0019) (-0.0005)		
BLANK TEST	0.000	14:56

Average = 0.0788
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 44.6°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.72 V	PASSED
(1.01 V - 2.37 V)	
Cooler Voltage = 1.64 V	PASSED
(1.03 V - 2.39 V)	
Bias Voltage = 78 V	PASSED
(47 V - 111 V)	
Chopper Freq = 530 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 28.9 in	

PUMP INFO

Flow Rate = 4.644 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

Xq = 0.098	0.27%	(0.00% - 4.00%)	PASSED
------------	-------	-----------------	--------