

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/12/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.01: 0.078
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 1237 psi

BLANK TEST	0.000	13:35
INTERNAL STANDARD	VERIFIED	13:35
EXTERNAL STANDARD	0.079	13:35
BLANK TEST	0.000	13:36
EXTERNAL STANDARD	0.079	13:36
BLANK TEST	0.000	13:37
EXTERNAL STANDARD	0.079	13:37
BLANK TEST	0.000	13:38
EXTERNAL STANDARD	0.079	13:39
BLANK TEST	0.000	13:40
EXTERNAL STANDARD	0.079	13:40
BLANK TEST	0.000	13:41

Average = 0.0790
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.6°C PASSED
Breath Tube = 43.2°C PASSED

PUMP INFO

Flow Rate = 4.596 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Charles R. Foster
Scientific Director
State Breath Alcohol Program

5/3/18

Subscribed and sworn before me this 03 day of 05, 2018


Carlie K. Bailey, Notary Public
My Commission Expires With Office



0034/25/18