

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100385

Received From AST Ketchikan

Date 04/10/2019

Analyst Derek Walton

1. Reason for instrument return:

Unable to upload or print tests.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload Unable to upload via modem line. Output data to USB and imported to DM Host.

5. Update agency information and polling status in DM Host Calling Program

DW 04/10/19

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)