

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100675

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/19/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.11: 0.081
LOT #: 09214080A2
EXPIRATION: 05/01/2016
TANK PRESSURE: 1133 psi

BLANK TEST	0.000	22:44
INTERNAL STANDARD	VERIFIED	22:44
EXTERNAL STANDARD	0.080	22:44
BLANK TEST	0.000	22:45
EXTERNAL STANDARD	0.080	22:45
BLANK TEST	0.000	22:46
EXTERNAL STANDARD	0.080	22:46
BLANK TEST	0.000	22:47
EXTERNAL STANDARD	0.080	22:47
BLANK TEST	0.000	22:48
EXTERNAL STANDARD	0.080	22:49
BLANK TEST	0.000	22:49

Average = 0.0800
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 3.01
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 48.0°C PASSED

PUMP INFO

Flow Rate = 5.200 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



Nita J. Bolz
Scientific Director
State Breath Alcohol Program

6/11/2015

Subscribed and sworn before me this 11th day of June 20 15


Nikki Roth, Notary Public
My Commission Expires With Office




DMB
6/10/15