

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100671

Received From AST-Fairbanks

Date 1/24/17

Analyst Brandi Barnett

1. Reason for instrument return:

fatal + undefined errors

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload unable to upload files, unable to view files.

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs unable to access reports screen or upload files.
error transferring files.

From Intoximeters

7. Intoximeters Repair Paperwork (attached)