

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 12/27/2017 Instrument Serial Number: 100369

Instrument Location: K'esugi Ken State Park , Chulitna

Supervisor Name and Agency: W. Stuart Kirksey Alaska State Troopers

Describe the Status Change or Issue with the Instrument:

Returned to Service

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 12/27/2017 / 2055 hours

Out of Service Date/Time: 12/12/17 12:00 BMB  
3/2/18

For Use by SCDL

Additional Notes

Instrument did not complete 12/12 voc. W. Kirksey attempted to troubleshoot on 12/21 but was unable to get VOC to pass. After I spoke to him I indicated he should attempt to purge the sample chamber to clear any environmental issues. This was done on 12/27/17 and the instrument was placed back in service.

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740