

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100406

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/23/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.57: 0.079
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 324 psi

BLANK TEST	0.000	16:24
INTERNAL STANDARD	VERIFIED	16:24
EXTERNAL STANDARD	0.079	16:25
BLANK TEST	0.000	16:25
EXTERNAL STANDARD	0.079	16:26
BLANK TEST	0.000	16:27
EXTERNAL STANDARD	0.079	16:27
BLANK TEST	0.000	16:28
EXTERNAL STANDARD	0.079	16:28
BLANK TEST	0.000	16:29
EXTERNAL STANDARD	0.079	16:29
BLANK TEST	0.000	16:30

Average = 0.0790
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 42.9°C PASSED

PUMP INFO

Flow Rate = 5.018 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 12/7/18

Charles R. Foster
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7 day of Dec 20 18


Kurt R. Palmer Sr., Notary Public
My Commission Expires With Office




OMB
11/27/18