

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100690

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/12/2022

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.05: 0.078  
LOT #: AG113402  
EXPIRATION: 05/14/2023  
TANK PRESSURE: 474 psi

BLANK TEST                      0.000    12:02  
INTERNAL STANDARD              ERROR    12:02

Internal Standard Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C              PASSED  
Breath Tube        = 46.0°C              PASSED

### PUMP INFO

Flow Rate = 4.056 L/M                  PASSED

### DETECTOR INFO

PUMP ON                                  PASSED  
PUMP OFF                                 PASSED

### FILTER INFO

Filter 1                                  PASSED  
Filter 2                                  PASSED  
Filter 3                                  PASSED

INTERNAL STANDARD                      PASSED

File Only DM 02/10/23