

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/21/2019

External Standard Test Values

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
Breath Tube = 46.1°C PASSED

PUMP INFO

Flow Rate = 4.818 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

Detector Overflow

File only. Instrument taken out of service on 02/21/2019 for multiple status messages. DMV 03/18/19