

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100355

Received From: Intoximeters

Date: 2/20/24

Analyst: Carrie Blaister

Reason for instrument return:

Filter wheel errors. Repaired by manufacturer.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

N/A
instrument
was not
in service.

From Intoximeters

Intoximeters Repair Paperwork (attached)

* No keyboard returned w/ instrument.

SRO Number: SRO-060459

Received Date: 01/17/2024

Completion Date: 01/23/2024

SRO Type: REPAIR

SRO Description: DMT Repair

Customer ID: C000003651

Customer PO: 01172024-OBryant

Contact: Colleen OBryant

Email: Colleen.obryant@alaska.gov

Ship To:

Alaska Scientific Crime Detection Laboratory
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Units on SRO

1	LEGDMT	LEGACY DMT
---	--------	------------

Service Line: 1		**All Instruments are Calibrated to Factory Specifications**	
Unit Type:	(LEGDMT) LEGACY DMT		
Serial Number:	50100355		
Reason:	DMT Status Messages		
	FltWl er		
Reason Notes:	Customer Reported Issue: Filter wheel errors		
Resolution:	Replaced		
	RPL comp		
Resolution Notes:	Replaced filter wheel solenoid. Replaced filter wheel chopper motor. Rebuilt filter wheel assembly. Performed fw test = 35:0 pass		
Parts:	<u>Qty</u>	<u>Part Number</u>	<u>Part Description</u>
	1	59-0080-00	DMT ELEC ASSY SOLENOID FILTR WHEEL 41344
Reason:	Preventive Maintenance		
	planned		
Reason Notes:	No TB-30 mod present.		
Resolution:	Update		
	DMTVR3		
Resolution Notes:	Installed.		
	1	001199-00	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN
Reason:	Preventive Maintenance		
	planned		
Reason Notes:	Preventative Maintenance		
Resolution:	Adjusted		
	Elec adj		
Resolution Notes:	Performed inspection and adjust instrument to factory specifications.		
	1	59-0006-00	DMT HRDW ASSY MOTOR 12X20.6MM 41310

Order Packing Slip Report

Intoximeters
Experience • Service • Integrity

2/5/2024 10:35:01 AM

Packing Slip: 299815

Pack Date: 2/5/2024

SRO: SRO-060459



From: Intoximeters, Inc
2081 Craig Road
St Louis MO 63146
USA

Customer: C000003651

Ship To: 1

Alaska Department of Public Safety
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Alaska Scientific Crime Detection
Laboratory
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Customer PO: 01172024-OBryant

Ship Via: FDX

FED EX

Weight: 36.00

Delivery Terms : 2D

Shipping Terms : ORIGIN

Packages: 1

Customer Ship Account :

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100355
1	001199-00	EA		1.00	0.00
	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN				
2	59-0006-00	EA		1.00	0.00
	DMT HRDW ASSY MOTOR 12X20.6MM 41310				
3	59-0080-00	EA		1.00	0.00
	DMT ELEC ASSY SOLENOID FILTR WHEEL 41344				