

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 2-15-19 Instrument Serial Number: B-100685

Instrument Location: Kodiak Police Dept.

Supervisor Name and Agency: Thomas Harris - Kodiak Police Dept.

**Describe the Status Change or Issue with the Instrument:**

Exchanging DMT B-100685 due to "Communication Error" during VOC process after performing tank change. Reset DMT as directed by Crime Lab staff. New DMT (B-100380) arrived and is ready for patrol use.

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 2/15/2019 at 1500 hours

For Use by SCDL	Additional Notes
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)  
For questions contact the Breath Alcohol Program at 907-269-5740