


Lot 2 085/19/10



Corporate Office:
P.O. Box 790, 7 Eastgate Drive
Jacksonville, IL 62651
217-245-2183
Fax: 217-243-7634
www.ilmoproducts.com



Certificate of Analysis

Product Code: **BAC105L080T**

0.080% BrAC

<u>Lot Number</u>	<u>Component</u>	<u>Concentration</u>	<u>Accuracy</u>	<u>Analytical Method</u>
33609080A1	Ethanol	209 ppm	+/- 0.002 BrAC	NDIR
	Nitrogen	Balance		

Balance Gas: Nitrogen

Expiration Date: 1/1/2012

Contents: 105 Liters @ 1025 PSIG

*NIST Standard Reference Material Certification of NTRM Batch No. 091602
Nominal 210 µmol/mol Ethanol in Nitrogen
for ILMO Products Co., Jacksonville, IL

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130128

MAY 19, 2010

OPERATOR'S NAME:
LOT33609080A1
OPERATOR'S NUMBER: EXP1/1/12
ALCO TARGET VALUE: .000
ALCO S/N: 2A

--- SUPERVISOR MODE ---

.000 ADJUSTED FOR 29.57 in		
ALCO TARGET	.079	10:02
BLANK TEST	.000	10:03
INTERNAL STANDARD	VERIFIED	10:03
ALCO TV 29.57 in	.078	10:04
BLANK TEST	.000	10:04
ALCO TV 29.57 in	.077	10:05
BLANK TEST	.000	10:05
ALCO TV 29.57 in	.077	10:06
BLANK TEST	.000	10:06
ALCO TV 29.57 in	.077	10:07
BLANK TEST	.000	10:07
ALCO TV 29.57 in	.077	10:08
BLANK TEST	.000	10:08

N = 5
SIN. = .079
AUG. = .0772
S.D. = .0004
S.E. = -2.3768

Travis Nelson

12-4-2009

Specialty Gas Lab Manager
ILMO Products Company

Date

The information contained herein is accurate and complete within the analytical methods employed. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to replacement of defective material.